



CERTIFICATE OF ANALYSIS
CITRANOX
FOR BECKMAN COULTER (BCI)

Date Of Manufacture: 11/3/2016
Manufactured by: Alconox, Inc.
Manufacturer Lot number: C9K6
BCI Lot number: 630003
Manuf. Cat. No.: 1801B
BCI Part No: 81912
Shelf Life Expiration Date: 10/2018 To meet BC requirements
Storage Conditions: Closed upright and in a cool dry place (15 - 30 deg C or 60 - 90 deg F).

<u>Value</u>	<u>-</u>	<u>Result</u>	<u>Acceptance Range</u>
Appearance:		OK	Clear pale olive colored liquid
pH: (1% solution in DI water)		2.8	2.0-3.0
Surface Tension: (1% solution Dyne/cm)		31.0	29.0 – 33.0
Specific Gravity		1.112	1.090 – 1.140
Number of Bar Code Labels Used for Labels		720	Not Applicable
Number of Bar Code Labels Used for Records		4	Not Applicable
Number of Bar Code Labels Destroyed		38	Not Applicable

Malcolm McLaughlin, QA

REF 81912

2018-10-31

LOT 630003



+H628819121A



+552103118630003AS



(01)15099590223571

(17)181031

(10)630003