



Alconox, Inc.

Critical Cleaning Experts

30 Glenn St., Suite 309, White Plains, NY 10603 USA

Tel.914.948.4040 - Fax.914.948.4088

www.alconox.com - cleaning@alconox.com

Factory: 61 Cornelison Ave. Jersey City, NJ 07304 USA

Date COA Created: 3/8/2019

CERTIFICATE OF ANALYSIS **CITRAJET**

Manufactured by: Alconox, Inc.

Sub lot No: H1C9 H1C9F, H1C9G, H1C9J, H1C9K, H1C9L, H1C9M

Sub lot numbers of Master lot No: H1L7

Manuf. Cat. no.: 2001, 2001-1, 2005, 2015, 2030, 2055

Date Of Manufacture: 3/7/2019

Expires end of: 03/21

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

<u>ANALYSIS</u>	<u>RESULT</u>	<u>ACCEPTED RANGE</u>
Appearance:	OK	Clear, olive colored liquid
pH: (1% solution in DI water)	3.1	2.1 - 3.4
Surface Tension:(1% solution Dyne/cm)	34.5	29.0 - 35.0
Specific Gravity:	1.13	1.09 - 1.15
Foam 1% Solution (110 Deg F)	NIL	NIL

Laura Falla, QA

Alconox, Inc.

Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.