



Alconox, Inc.

The Leader in Critical-Cleaning

30 Glenn St., Suite 309, White Plains, NY 10603 USA

Tel.914.948.4040 • Fax.914.948.4088

www.alconox.com • cleaning@alconox.com

Factory: 61 Cornelison Ave. Jersey City, NJ 07304 USA

Date COA Created: 8/6/2018

CERTIFICATE OF ANALYSIS
KEYLAJET

Manufactured by: Alconox, Inc.

Sub lot No: K1H8

Sub lot numbers of Master lot No: None

Manuf. Cat. no.: 2405, 2415, 2455

Date Of Manufacture: 8/1/2018

Expires end of: 08/20

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

ANALYSIS

RESULT

ACCEPTED RANGE

Appearance:	OK	Clear pale yellow liquid
pH: (1% solution in DI water)	12.9	11.5 - 13.5
Surface Tension:(1% solution Dyne/cm)	31.4	30.0-34.0
Specific Gravity:	1.317	1.25-1.39

Malcolm McLaughlin, QA

Alconox, Inc.

Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.