



Alconox, Inc.

Critical Cleaning Experts

30 Glenn St., Suite 309, White Plains, NY 10603 USA

Tel.914.948.4040 - Fax.914.948.4088

www.alconox.com - cleaning@alconox.com

Factory: 61 Cornelison Ave. Jersey City, NJ 07304 USA

Date COA Created: 10/7/2019

CERTIFICATE OF ANALYSIS
KEYLAJET

Manufactured by: Alconox, Inc.

Sub lot No: K1J9

Sub lot numbers of Master lot No: None

Manuf. Cat. no.: 2405, 2415, 2455

Date Of Manufacture: 9/26/2019

Expires end of: 09/21

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

<u>ANALYSIS</u>	<u>RESULT</u>	<u>ACCEPTED RANGE</u>
Appearance:	OK	Clear pale yellow liquid
pH: (1% solution in DI water)	12.7	11.5 - 13.5
Surface Tension:(1% solution Dyne/cm)	32.7	30.0-34.0
Specific Gravity:	1.33	1.25-1.39

Laura Falla, QA

Alconox, Inc.

Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.