



Alconox, Inc.

The Leader in Critical-Cleaning
30 Glenn St., Suite 309, White Plains, NY 10603 USA
Tel.914.948.4040 • Fax.914.948.4088
www.alconox.com • cleaning@alconox.com

Factory: 61 Cornelison Ave. Jersey City, NJ 07304 USA

Date COA Created: 12/5/2016

CERTIFICATE OF ANALYSIS

SOLUJET

Manufactured by: Alconox, Inc.

Sub lot No: S1L6 S1L6F, S1L6G, S1L6J, S1L6K, S1L6L, S1L6M

Sub lot numbers of Master lot No: S1A5

Manuf. Cat. no.: 2101, 2101-1, 2105, 2115, 2130, 2155

Date Of Manufacture: 11/28/2016

Expires end of: 11/18

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

| <u>ANALYSIS</u> | <u>RESULT</u> | <u>ACCEPTED RANGE</u> |
|---------------------------------------|----------------------|------------------------------|
| Appearance: | OK | pale straw colored liquid |
| pH: (1% solution in DI water) | 12.2 | 11.5 - 13.0 |
| Surface Tension:(1% solution Dyne/cm) | 31.9 | 31.5 - 36.5 |
| Specific Gravity: | 1.282 | 1.26 - 1.34 |

Malcolm McLaughlin, QA

Alconox, Inc.

Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.