



Alconox Inc.

Critical Cleaning Experts

30 Glenn St., Suite 309, White Plains, NY 10603 USA

Tel.914.948.4040 • Fax.914.948.4088

www.alconox.com • cleaning@alconox.com

Date COA Created: 4/12/2023

## CERTIFICATE OF ANALYSIS ALCOJET

**Manufactured by: Alconox Inc.**

Sub lot No: 04-06-23-12

Sub lot numbers of Master lot No: J1A2

Manuf. Cat. no.: 1401, 1403, 1404, 1404-1, 1425, 1450

Date Of Manufacture: 4/7/2023

Expires end of: 04/25

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

**ANALYSIS**

**RESULT**

**ACCEPTED RANGE**

|   |      |              |
|---|------|--------------|
| Appearance:   | OK   | White powder |
| Solubility: (1% solution at 25 deg C, with agitation) |      | Complete     |
| pH:(1% solution in DI water)                          | 11.3 | 11.0 - 12.5  |
| Surface Tension:(1% solution Dyne/cm)                 | 32.5 | 31.0 - 36.0  |
| Foam (1% in DI water)                                 | NIL  | NIL          |

Laura Falla, QA

Alconox, Inc.

*Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.*