



Alconox, LLC

Critical Cleaning Experts

30 Glenn St., Suite 309, White Plains, NY 10603 USA

Tel.914.948.4040 • Fax.914.948.4088

www.alconox.com • cleaning@alconox.com

Date COA Created: 10/8/2025

CERTIFICATE OF ANALYSIS **CITRAJET**

Manufactured by: Alconox, LLC

Lot No: 09-09-25-01

IRT Lot No: 09-05-23-02

Manuf. Cat. no.: 2001, 2001-1, 2005, 2015, 2030, 2055

Date Of Manufacture: 9/9/2025

Expires end of: 03/29

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

ANALYSIS

RESULT

ACCEPTED RANGE

Appearance:	OK	Clear, olive colored liquid
pH: (1% solution in DI water)	3.1	2.1 - 3.4
APHA color:	30	0 - 150
Specific Gravity:	1.13	1.09 - 1.15
Foam 1% Solution (110 Deg F)	NIL	NIL

Laura Falla, QA

Alconox, LLC

Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.