



Alconox, LLC

Critical Cleaning Experts

30 Glenn St., Suite 309, White Plains, NY 10603 USA

Tel.914.948.4040 • Fax.914.948.4088

www.alconox.com • cleaning@alconox.com

Date COA Created: 3/17/2026

CERTIFICATE OF ANALYSIS
ALCOJET

Manufactured by: Alconox, LLC

Lot No: 03-12-26-12

IRT Lot No: 07-01-24-12

Manuf. Cat. no.: 1401, 1403, 1404, 1404-1, 1425, 1450

Date Of Manufacture: 3/12/2026

Expires end of: 09/29

Storage Conditions: Closed upright and in a cool dry place (should be 5 - 40 deg C or 41 - 104 deg F).

ANALYSIS

RESULT

ACCEPTED RANGE

Appearance:

OK

White powder

Solubility: (1% solution at 25 deg C,
with agitation)

COMP

Complete

pH:(1% solution in DI water)

11.5

11.0 - 12.5

Free alkalinity: (meq/g)

4.9

3.7 - 5.5

Foam (1% in DI water)

NIL

NIL

Laura Falla, QA

Alconox, LLC

***Contains no class 1, class 2, class 3, or class 4 residual
solvent ingredients as defined by ICH Q3C guidelines.***