



Alconox, LLC

Critical Cleaning Experts

30 Glenn St., Suite 309, White Plains, NY 10603 USA

Tel.914.948.4040 • Fax.914.948.4088

www.alconox.com • cleaning@alconox.com

Date COA Created: 5/25/2026

CERTIFICATE OF ANALYSIS
SOLUJET

Manufactured by: Alconox, LLC

Lot No: 05-21-26-01

IRT Lot No: 03-12-25-01

Manuf. Cat. no.: 2101, 2101-1, 2105, 2115, 2130, 2155

Date Of Manufacture: 5/21/2026

Expires end of: 11 / 29

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

<u>ANALYSIS</u>	<u>RESULT</u>	<u>ACCEPTED RANGE</u>
Appearance:	OK	pale straw colored liquid
pH: (1% solution in DI water)	12.3	11.5 - 13.0
APHA color:	90	0 - 200
Specific Gravity:	1.28	1.26 - 1.34

Laura Falla, QA

Alconox, LLC

Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.