



Alconox, LLC

Critical Cleaning Experts

30 Glenn St., Suite 309, White Plains, NY 10603 USA

Tel.914.948.4040 • Fax.914.948.4088

www.alconox.com • cleaning@alconox.com

Date COA Created: 9/3/2026

CERTIFICATE OF ANALYSIS
CITRANOX

Manufactured by: Alconox, LLC

Lot No: 08-28-26-01

IRT Lot No: 01-21-25-01

Manuf. Cat. no.: 1801, 1801-1, 1805, 1815, 1832, 1832-1, 1855

Date Of Manufacture: 8/28/2026

Expires end of: 02 / 30

Storage Conditions: Closed upright and in a cool dry place (should be 15 - 30 deg C or 60 - 90 deg F).

ANALYSIS

RESULT

ACCEPTED RANGE

Appearance:	OK	Clear, olive colored liquid
pH: (1% solution in DI water)	2.8	2.0 - 3.0
APHA color:	70	0 - 200
Specific Gravity:	1.11	1.09 - 1.14

Laura Falla, QA

Alconox, LLC

Contains no class 1, class 2, class 3, or class 4 residual solvent ingredients as defined by ICH Q3C guidelines.